

Glider Prep and Deployment Checklists

Glider -
deployment

wd-476-203011371454

Project

Moffat Antarctica
Career

Deployment
dates

1/13/23 - 1/22/23

Location/notes

Near Adelaide Island,
Antarctica

- leave from aft, likely
at ~~1000~~ fin connection

Extant

Notes

1) Glider check-out sheet

2) Ballasting/dunk sheets

3) Deployment checklists
(on boat, shore side)

4) Glider check-in sheet

- glider did not return to
RU after deployment -
went straight to UDel,
no RU check in

5) Misc. (science, etc.)

 CTD

 Optode

 LISST

6) Other

 Compass

✓ ACD compass

GLIDER	UD-476
PREPARER	Nicole
PREP DATE	8/23/2022
LOCATION / MISSION	MOFFAT / ANTARCTICA
DENSITY @ TEMP	1026.0 @ 0°C
INSURED?	

		Calibration Date (user/factory)
SCIENCE BAY SERIAL NUMBERS	1) CTD 9537	11/12/2020
	2) PLBB ^{SLL} 5499	11/14/2020
	3) PAR 50283	11/17/2020
	4) Oxy 4831 sln 334	Factory 10/12/2020 in-house cal 8/24/2022
	5) AOD 11619, head 6483	
	6)	

PRE-SEAL TAKE PICTURES OF CONNECTORS AT EACH SEALING JOINT

✓

FORE CHECK

Check pump & pitch threaded rods (clean and grease) ✓ Leak detect in place, batteries secure, grab & wiggle pitch battery to check secure, white guides free, no metal shavings, bottles installed ✓

Grounded nose? ✓

Dessicant Exposed? ✓

PAYLOAD CHECK

Special Sensors / Additional Sensors? Aquadepp CTD cable clear, no leak at CTD joint, no leak at pucks ✓

Grounded? ✓ Fore Sci Ring No CTD ✓

Corrosion? ✓ Aft Sci Ring No Other? ✓

AFT CHECK

Iridium Card Installed (SIM #) (if not standard) ✓ DoD 8988 169 763 060 258164

Flash Card Check (remove old files, backed up? See Software section) ✓

Inspect strain on connectors/worn connectors ✓

Battery secured ✓

Ballast bottle present ✓

Aft cap clean/clear of leak ✓

Ejection weight stem grounded? Should it be? (Version specific) No

Thruster greased? N/A

Ensure safety of ballast pump prior to powering glider

Battery check: G2/G1 turn glider on with only 1 battery connected; G3 use BMS current

Aft Pack Voltage ✓ Turned on

Pitch Pack Voltage ✓ Turned on

Nose Packs Voltage ✓

Emer (if possible) Voltage ✓

Cabling/connectors - lithium vs. alkaline circuit correct? ✓

POST-SEAL, pre-ballast

GENERAL

Pick Point Present? No Special Cargo? No

HARDWARE

Nose cone and pump bladder inspection ✓

Anode grounded? ✓ Anode size / remainder New

Pressure Sensor Check (corrosion, clear) ✓

Aft sensor ✓ Payload sensor ✓

Ejection weight assembly ok/not seized? ✓

* Do a logging on for all these checks, take note of log and transfer before deployment

SENSOR RETURN 0194.0000 // 234-0-0

put c_science_send_all 1 ✓
 put c_science_all_on 8 ✓
 put c_science_on 3 ✓
 All sensors reporting values? ✓

CTD 0193.0000 // 233-0-0

Tank static comparison OK? ✓
 Pumped CTD operational? ✓
 Plot ballast *BD log, sci_water_pressure non-noisy and near < .5 m ✓

OPTODE

Check in completed? ✓
 Saturation reading in air 101.601

OPTICS 0195.0000 // 235-0-0

Check max return using fluoro sticks ✓ 4130 both
 Check dark counts with sensor covered bb=48 chl=42
 Optics file name 0195.0000

LISST

Clean LISST and perform ZSCAT N/A

OTHER

Aquadopp Set-up
 Agd Compass cal 8/24/2022

OUTSIDE

GPS Almanac/firmware updated? ✓
 GPS check Latitude 40 28.749 N Longitude -74 26.222 E
 Iridium connect ✓
 zero_poser_pressure ✓
 Air bladder shutoff (time)? ✓ 2:38
 Compass calibration ✓
 For deep gliders, put c_de_oil_vol -1000 to fully retract oil inside reservoir ✓

log exp. 0 put on slider for DOD sim

ADDITIONAL

***WARNING: Advanced knowledge required to avoid damage/injury

Check burn wire - disconnect, then put c_weight_drop 1, confirm 12 V ✓

Fore leakdetect _____ Science _____ Aft leakdetect _____

THRUSTER

Report ++ m_thruster_current
 Put c_thruster_on 20
 Verify thruster spins clockwise and current value updates regularly
 Put c_thruster_on 0 to turn off N/A

POWERED

Put m_coulomb_amphr_total accordingly (0 = new batteries)
Put f_coulomb_battery_capacity (Alk=155, Lilon=200, li=450,625)
Vacuum @ T @ ballast 5.89 @ 24.6C Stabilized m_battery
Get m_tot_num_inflections. Verify relative < 20000 or sufficient
Get m_leakdetect_voltage, science, forward (>2.3)
Get m_digifin_leakdetect_reading (less than 1019 requires service)
Altimeter test - put c_alt_time 0, verify chirp, note m_altimeter_voltage
Verify Argos ping ✓ Wiggle for 5 minutes

5.93 @
23.3C
Ballast+

✓
✓ 450
14.97 volts
✓ 9813
✓ 2.47, 2.49
✓ 1023
✓ 2.465 volts
✓

SOFTWARE

(paths are RU specific)

GENERAL

Backup Glider and Science Cards
COOL/gliderData/glider_OS_backups/"glider name"
Format both CF cards - FAT Format
Apply new copy of latest TWR Software Image
For Glider: COOL/gliderData/gliderDos_releases/archived/"version"/target-glider
For Science: COOL/gliderData/gliderDos_releases/archived/"version"/target-science
Copy/overwrite STATE and CONFIG Folders
FW Transfer latest RU Software Image *cur put into to-glider on dockserver*
COOL/Glidrs/Glider Software Image/"use most recent image"
Software Version 8.3 Configure TBDlist
Date OK? ✓ Configure NBDlist

✓
✓
✓
✓
✓
✓
✓
✓

CONFIG

simul.sim deleted

✓

IMAFILES

goto_i10.ma (set x_last...)
yo*.ma, surfac*.ma pertinent for each glider and test missions

○
✓

MISSIONS

b_arg: undervolts: 10.75V alkaline, 9-10 V Li3S, 13.5V Li4S, 12.75V Lilon
Remove unused sample behaviors in missions

✓
✓

AUTOEXEC.MI

Iridium: Numbers may vary. Listed: Main - Rutgers Alternate - TWR
Irid Main: 88160000592 88160000500 Irid Alt: 17818711614
u_iridium_failover_retries = 10 ✓ Ver 7.15 u_iridium_idle -1?
sci timestamp sensors (ctd41cp) ✓ Calibration coefficients
Reset the glider, observe any errors ✓ get f_max_working_depth

1577818711614 DOD #5
206

CACHE MANAGEMENT

del ..\state\cache*.*
after *bdlist.dat are set (exit reset):
logging on; logging off
send ..\state\cache*.cac
send *.mbd *.sbd *.tbd *.nbd

✓
✓
✓
✓
✓

DOCKSERVER

Version
Check script

8.5.0-1

TWR BACKUP

Confirm to-glider folder clear
Confirm correct script running

○

* **Software Burning Tips** : if using Procomm or local folder, copy all the files from the software image locally. Then proceed to edit them for the glider and do a mass freewave transfer of the files. Save these files or prepare the to-glider with these files

NOTES

burnapp to 8.3

- removed m-pump-stress-remaining-cycles + m-pump-stress-track
from longterm.dat

tested strobe - works

Adjusted Pitch-motor cal-b value, battery was a little far back,
shifted battpos forward 0.38 inches

Sci cal values verified (PAR, FLBB)

pump cal verified

★ Need to check on pressure sensor cal values!

- using typed min + max values on cal sheet - confirmed these values are in
autobexc.mi
↳ not hand written values

CTD pump works - but sounds a little strange in lab.

AQD head pointed forward

AQD feet: 1 sec rate
App A checklist
compass test, pressure test
ping in air: 20-30 counts sig
ping in water → sig strength up
on-line → range check
pressure offset in tank

UD 476

8/19/2022

Pitch motor

- moved to ballast \emptyset fine
- moved 0 to 0.2 ~~no~~ fine
- moved 0.2 to 0.5 fine
- tried 0.5 to 1 and didn't move, ^{pitch} motor got taken out of service.
- tried to put pitch-motor back into service a couple times and it wouldn't go.

Front $\frac{1}{2}$ inc' back $\frac{1}{4}$ " thread left

M1 .89

M2 -.88

glider 1.74 1.77 ✓

m shcher : 1.962
b -1.743

Changing b to -1.743 changed batt pos by .38
- more negative b = ~~more~~ ^{shift} forward battery

		MASS (g)	COMMENTS
Deployment			
Moffat Antarctica			
Glider	GLIDER		Adjust trim
UD_476	FORE STEM (minus FBB1,2)		removed 44g from aft ballast bottle
	FORE HULL		
	AFT STEM (red plug, card)		
	AFT HULL		
	COWLING		add 44g to STBD Front ballast bottle (try to fix roll)
	SCREWS (vacuum, cowling, aft battery)		
Date	PAYLOAD BAY		
8/25/2022	WINGS		
	OTHER - Aquadopp		
Preparer	BATTERIES		
Nicole	AFT BATTERY		
	PITCH BATTERY		
	FORE BATTERY 1, 2, EMER		
	AFT BOTTLE	24.3	
Final	FORE BOTTLE 1 (stbd) (FBB1)	325	
Dunk for Roll only	FORE BOTTLE 2 (port) (FBB2)	239.1	
	OTHER		

ENTIRE VEHICLE (Ohaus Scale)

Tank Specifics	Glider Specifics
Tank Density (kg/m ³)	Glider Volume (L)
1020.35	60.475
Tank Temperature (C)	Total Mass (kg)
22.10	0.588
Weight in Tank (g)	Glider Density (in air)
280.00	9.73
Target Specifics	Volume Change (temperature induced)
Target Density (kg/m ³)	Volume Change (target) (mL)
1026.00	-60.2
Target Temperature (C)	Coefficient of Thermal Expansion
0.00	4.50E-05
	Carbon hulls
	4.50E-05
	Aluminum hulls
	7.50E-05

Ballasting Using Volume	Ballasting Using Mass
Should Hang (in tank) (g)	Adjust Glider Mass (entered volume) (g)
280.0	61397.2
Adjust by (g)	Glider Density (target water, using mass)
0.0	9.7
Weight Change (no dunk) (g)	
Glider Density (target)	
1026.0	

H MOMENT (rad)	(deg)
Angle of Rotation (before)	-0
Angle of Rotation (after)	-2.6
Angle of Rotation	0.0
Weight on Spring (after)	-0
Weight added	-2.6
Radius of Hull	125
H-distance	0.0

125 for G2+, deeps

MISC MASSES & VOLUMES

Pick point - 40 mL - 107 g air - 66 g water

Wing Rail Weights - 1.8 mL @ 15.4 g each ~ 13.5 g in water

WMT Transceiver - 173 mL - 162 g water

FIRE Shroud SN02 (ru01) - 266 mL - 112 g water

Optode - 130 mL - 92 or 190 g (plastic or titanium)

LISST Bay - roughly 6.55 L

Carbon Fiber wing Air -> Water Ratio 0.437209

Plastic Wing Air -> Water Ratio 0.487603

2K syntactic foam 100 g gives you 155 buoyancy

MASS (g)

COMMENTS

Deployment		FORE STEM (minus FBB1, 2) ^{9926.4}	9926.4	wt 4-wt desiccant + battery	Desiccant = 125.25 m.k.2
Moffat Antarchica		FORE HULL	4558.4		
Glider		AFT STEM (red plug, card)	6713.4	w/optode	
UD-476		AFT HULL	4566.0		
		COVERLING	1225.0		
		SCREWS (vacuum, cowling, aft battery)	24.5		
		PAYLOAD BAY	1225.0	w/ battery mounting bracket	3 pins front wings rails secured 2 pins base
		WINGS	446.7		
		OTHER - Airframe + mounts	3453.9		aluminum battery + swap to 2x Lithium after 1st dunk -75g
8/19/2022		AFT BATTERY	10083.0		
Preparer		PITCH BATTERY	8960.4		
Nicole		FORE BATTERY 1, 2, EMER	2047.6	included in nose.	
		AFT BOTTLE	120.0		
		FORE BOTTLE 1 (std) (FBB1)	231.2		
		FORE BOTTLE 2 (port) (FBB2)	207.3		
		OTHER FBB3	4584.7		

ENTIRE VEHICLE (Chaus Scale)

Tank Specifics		Glider Specifics	
Tank Density (kg/m ³)	1023.00	Glider Volume (L)	57.000
Tank Temperature (C)	23.00	Total Mass (kg)	0.000
Weight in Tank (g)	0.00	Glider Density (in air)	0.00
Target Specifics		Volume Change (temperature induced)	
Target Density (kg/m ³)	1023.00	Volume Change (target) (mL)	0.0
Target Temperature (C)	23.00	Coefficient of Thermal Expansion	4.50E-05
		Carbon hulls	4.50E-05
		Aluminum hulls	7.50E-05
Glider Volume (at lab temp) (L)	0.000		

H-MOMENT (rad)	(deg)
Angle of Rotation (before)	0.0
Angle of Rotation (after)	0.0
Angle of Rotation	0
Weight on Spring (after)	0.0
Weight added	290
Radius of Hull	107
H-distance	####

125 for G2+, deeps

MISC MASSES & VOLUMES

- Pick point - 40 mL - 107 g air - 66 g water
- Wing Rail Weights - 1.8 mL @ 15.4 g each ~ 13.5 g in water
- VMT Transceiver - 173 mL - 162 g water
- FIRE Shroud SN02 (ru01) - 266 mL - 112 g water
- Optode - 130 mL - 92 or 190 g (plastic or titanium)
- LISST Bay - roughly 6.55 L
- Carbon Fiber wing Air -> Water Ratio 0.437209
- Plastic Wing Air -> Water Ratio 0.487603
- 2k syntactic foam 100 g gives you 155 buoyancy

Ballasting Using Volume

Should Hang (in tank) (g)	0.0
Adjust by (g)	0.0
Weight Change (no dunk) (g)	
Glider Density (target)	1023.0

Ballasting Using Mass

Adjust Glider Mass (entered volume) (g)	58311.0
Glider Density (target water, using mass)	0.0

0487603
Ballasting Sheet

GLIDER: UD 476

Iteration 1 Log File 230-0-0 Date / Location RU / 8/19/2022

0192.0000

476

Ballast

Notes

Ver 8.2

FORE

EB

SB

AFT

FBB1 stbd 231.2

• = 168g

Front Scale

Aft Scale

FBB2 Port 207.3

pitch motor taken out of service...

78

506

Aft BB 120

Instrument: 1581

Instrument: _____

T = 22.017

T = _____

Roll _____

C = 4.311

C = _____

Ballast 1.076cc

D = 1020.17

D = _____

Battery 0.008 in

Iteration 2 Log File 233-0-0 Date / Location 8/22/2022 RU

0193.0000

UD

Ballast

Notes

FORE

EB

SB

AFT

FBB1 stbd 280.0

F A

Front Scale

Aft Scale

FBB2 Port 257.3

-90 506

178

110

Aft BB 57.3

+75 -75 ptd batt wrong

Instrument: 1581

Instrument: glider

-3.1°

-15 431

VS

T = 21.946

T = 21.948

Roll -0.054 rad

+223 -223 trim

C = 4.335

C = 4.331

Ballast 1.065cc

-60 -60 ballast

D = 1020.35

D = _____

Battery 0.020 in

+163 283

+20 -20

+183 -303

Done 2 notes continue below

Iteration _____ Log File _____ Date / Location _____

Ballast

Notes

FORE

EB

SB

AFT

FBB1 stbd _____

+183 -303 2nd -120

Front Scale

Aft Scale

FBB2 Port _____

-160 -240 -2 SBA pic

Done 2 H-Moment

scale w/ mass: 516

Aft BB _____

+343 -63

Instrument: _____

Instrument: _____

+343 -63 + -400 = -120

T = _____

T = roll w/ mass: -0.387 rad

Roll _____

Add 318g metal rod front, bottom

C = _____

mass: C = 294g

Ballast _____

Change AFD batteries to lithium -75.2g

D = _____

D = _____

Battery _____

↳ add 100.2g to front BBs
↳ 50g each side

H = 4.8

OHAUS Scale: 14.1
Glider + cart = 76.10kg

GLIDER: UD-476

Iteration 3 Log File 235-1-0 ^{0196.0000} Date / Location Ru 8/24/2022

Ballast				Notes	
FORE	EB	SB	AFT	F	A
FBB1 stbd <u>280.0 (NC)</u>				<u>178</u>	<u>110</u>
Front Scale				<u>+34</u>	<u>+34</u> <i>Balku trim</i>
Aft Scale				<u>+20</u>	<u>-20</u> <i>make nose heavy</i>
<u>158</u>				<u>-4</u>	<u>-4</u> <i>remove P5</i>
Instrument: <u>1581</u>				<u>-18</u>	<u>+10</u>
T = <u>22.145</u>				Roll <u>0.042 rad = 2.4°</u>	
C = <u>4.317</u>				Ballast <u>1.065cc</u>	
D = <u>1020.10</u>				Battery <u>0.014 in</u>	

Iteration 4 Log File _____ Date / Location Ru 8/25/2022

Ballast				Notes	
FORE	EB	SB	AFT	F	A
FBB1 stbd <u>280.0 (NC)</u>				<i>Add pickpoint aft SB</i>	
Front Scale				<i>remove 2 new part, 3 new stbd</i>	
Aft Scale				<i>reverse direction of A/D</i>	
<u>116</u>				<i>remove yellow core</i>	
Instrument: <u>1581</u>					
T = <u>22.104</u>				Roll <u>-0.068 rad = -3.9°</u>	
C = <u>4.354</u>				Ballast <u>0.026 in</u>	
D = <u>1020.35</u>				Battery <u>1.065cc</u>	

Iteration 5 Log File _____ Date / Location 8/25/2022

Ballast				Notes	
FORE	EB	SB	AFT	F	A
FBB1 stbd <u>325.0</u>				<u>116</u>	<u>104</u>
Front Scale				<u>+24</u>	<u>-24</u> <i>trim</i>
Aft Scale				<u>+20</u>	<u>-20</u> <i>40g nose heavy</i>
<u>No Ballast done only roll check</u>				<u>+44</u>	<u>-44</u>
Instrument: _____					
T = _____				Roll <u>-0.045 rad = -2.58°</u>	
C = _____				Ballast _____	
D = _____				Battery _____	

Glider / Mission:

UD476 AQD

Cal Location

DMCS cart

Date

8/25/2022

Operator


Nicole, Brian

** AQD compass re-calibrated with glider attached

LEVEL		
HAND	GLIDER	ERROR
0	1.8	1.8
45	50.9	-0.1
90	93.7	3.7
135	137.4	4.4
180	177.9	0.9
225	219.4	-0.6
270	260.2	2.2
315	310.5	3.5
		0
		0
		0

DOWNCAST		
HAND	GLIDER	ERROR
0		0
30		0
60		0
90		0
120		0
150		0
180		0
210		0
240		0
270		0
300		0
330		0

UPCAST		
HAND	GLIDER	ERROR
0		0
30		0
60		0
90		0
120		0
150		0
180		0
210		0
240		0
270		0
300		0
330		0



Slocum CTD Comparison Check

GLIDER: UD-476 SB: ^{PAR Bay} 9537p DEPLOYMENT: Moffat Antarctica

Pre-Deployment

Date: 8/22/2022

SBE19 s/n: <u>1581</u>	Glider: <u>9537p</u>
Temperature: <u>21.946</u>	Temperature: <u>21.948</u>
Conductivity: <u>4.335</u>	Conductivity: <u>4.331</u>

Notes:

8/22/22 could not feel glider CTD pump running in tank, but did observe ripples, so appeared to be working.

Post-Deployment

Date: _____

SBE19 s/n:	Glider:
Temperature:	Temperature:
Conductivity:	Conductivity:

Notes:

*** CTD Maintenance if comparison is not acceptable (reference SeaBird Application Note 2D)

1. Perform CTD backward/forward flush with 1% Triton X-100 solution
2. Perform CTD backward/forward flush with 500 – 1000 ppm bleach solution
3. Perform the same on a pumped unit, just different approach
4. Repeat comparison test if above results not within $T < .01$ C, $C < .005$ S/m

Oxygen Optode Check & Calibration *pre UD-476 Moffat Antarctica*

OPTODE SN	<u>334</u>	DATE	<u>8/24/2022</u>
FOIL ID	<u>1206EM</u>	AIR PRESSURE (hPa)	<u>29.82 mmHg = 1009.82 hPa = 757.428 mmHg</u>
PRE SALINITY	<u>35</u>	CALIBRATED?	<u>yes</u>

* REMEMBER TO ISSUE THE SAVE COMMAND AFTER CHANGING VALUES

pre-interval = 2

100% SOLUBILITY	<u>267.38 μM = 8.56 ppm</u>	TITRATION	<u>7.8 ppm, 7.8 ppm</u>
	* μM = ppm * 1000 / 32	EPA Na2S2O3 Check	<u>2.14</u> mL
		Sodium Sulfite / mL	<u>16g / 800mL</u> %

PRE-CHECK	
100%	0%
Conc (μM) = <u>251.953</u>	Conc (μM) = <u>1.484</u>
Saturation (%) = <u>95.842</u>	Saturation (%) = <u>0.545</u>
Temp (°C) = <u>23.978</u>	Temp (°C) = <u>22.147</u>
Phase = <u>30.407</u>	Phase = <u>60.426</u>

POST-CAL	
100%	0%
Conc (μM) = <u>251.953 251.953</u>	Conc (μM) = <u>-0.052</u>
Saturation (%) = <u>95.842 95.842</u>	Saturation (%) = <u>-0.019</u>
Temp (°C) = <u>23.978</u>	Temp (°C) = <u>22.476</u>
Phase = <u>30.407</u>	Phase = <u>60.442</u>

*262.004
99.455
23.866
30.649*

GLIDER CONFIG	
POST SALINITY	<u>34</u>
TEXT OUTPUT OFF	<input checked="" type="checkbox"/>

Post Interval = 2

* REMEMBER TO ISSUE THE SAVE COMMAND AFTER CHANGING VALUES

PRESSURE TRANSDUCER CALIBRATION (PSI)

Date: 10-Dec-2020
 Glider: 476


Dead Weight Kg/cm ²	Bar	AD1, 1 Volts
0	0	0.087
150	10.34	0.194
350	24.13	0.34
550	37.92	0.491
750	51.71	0.639
900	62.05	0.752
1150	79.29	0.934
1350	93.08	1.084
1550	106.87	1.201

m= 94.454 BAR's per Volt
 b= -8.341 BAR

VOLTS @ 0 BAR= 0.088
 VOLTS @ 140 BAR= 1.571

F ocean pressure min 0.087 V
max 1.201

A number: 58.298




PUMP MOTOR

Date: 12/21/20
FWD: 218

Pump position in inches	AD2,0 CC's	Pot. reading in Volts
0.0	-252	0.345
1.6	0	0.861
3.2	252	1.359

m=	496.989 CC's per Volt
b=	-424.926 CC's

Number: 20.252



Biospherical Instruments Inc.

CALIBRATION CERTIFICATE

Calibration Date 11/17/20
Model Number QSP2155 Firmware Version v 1.2
Serial Number 50283
Operator TPC
Standard Lamp V-040(11/12/20)
Sensor Operating Voltage Range: 6 to 15 VDC (+)
Output Polarity: Positive

A=125

Sensor Output Voltage:

Sensor Illuminated	<u>129.1</u> mV
Sensor Dark	<u>-9.9</u> mV
Sensor Net Response	<u>119.2</u> mV
RG780	<u>10.0</u> mV

Corrected Lamp Output:

Output In Air (same condition as calibration):

1.558E-02 uE/cm²sec

Output Corrected for Immersion in Water: Using Immersion Coefficient of:

2.751E-02 uE/cm²sec

0.5664 (Collector Type: SC-3)

Calibration Scale Factor:

(To calculate irradiance, divide the net voltage reading in Volts by this value.)

<u>Dry:</u>	7.646E+00 Volts/(uE/cm ² sec)
<u>Wet:</u>	4.331E+00 Volts/(uE/cm ² sec)
<u>Dry:</u>	7.646E-04 Volts/(uE/m ² sec)
<u>Wet:</u>	4.331E-04 Volts/(uE/m ² sec)

Notes:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. To approximate the sensor's saturating irradiance, multiply the calibration factor by the sensor power supply voltage, minus one volt.
4. The collector should be cleaned frequently with alcohol.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 11/14/2020

S/N: FLBBSLC-5499

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark counts})$$

Dark counts	Digital 38 counts
Scale Factor (SF)	0.0071 $\mu\text{g/l/count}$
Maximum Output	4130 counts
Resolution	1.0 counts
Ambient temperature during characterization	22.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $\text{SF} = x / (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518
620 Applegate St.
Philomath, OR 97370



754-271-2217
www.wetlabs.com

Scattering Meter Calibration Sheet

11/14/2020

Wavelength: 700

S/N FLBBSLC-5499

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• Scale Factor for	=	1.799E-06 (m ⁻¹ sr ⁻¹)/counts
• Output	=	meter reading - counts
• Dark Counts	=	31 counts
Instrument Resolution	=	1.0 counts 1.89E-06 (m ⁻¹ sr ⁻¹)

Definitions:

- **Scale Factor:** Calibration scale factor, $\beta(\theta_c)/\text{counts}$. Refer to User's Guide for derivation.
- **Output:** Measured signal output of the scattering meter.
- **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

WET Labs
14000 14th St
P.O. Box 1070
Boulder, CO 80501



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECD Chlorophyll Fluorometer Characterization Sheet

Date: 01/13/2020

S/N: FLBBSLC-5499

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{Chl. } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark counts})$$

Dark counts

Digital

38 counts

Scale Factor (SF)

0.0071 $\mu\text{g/l/count}$

Maximum Output

4130 counts

Resolution

1.0 counts

Temperature (°C) during characterization

22.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Calculated using the following equation: $\text{SF} = x / (\text{output} - \text{dark counts})$, where x is the concentration of the solution. For this instrument characterization, SF is used to derive instrument output concentration from the raw digital signal of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Minimum resolution of 1 count of collected data.

Chlorophyll concentration, biomass and chlorophyll-a concentrations (Chl-a) is highly variable. The scale factor listed on this fluorometer was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be representative of the natural population and was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using analytical laboratory measurement techniques on discrete samples. For additional information on determining chlorophyll concentration, see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

Scattering Meter Calibration Sheet

11/14/2020

Wavelength: 700

S/N FLBBSLC-5499

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• Scale Factor for 700 nm	=	1.799E-06	(m ⁻¹ sr ⁻¹)/counts
• Output	=	meter reading	counts
• Dark Counts	=	31	counts
Instrument Resolution	=	1.0	counts 1.89E-06 (m ⁻¹ sr ⁻¹)

Definitions:

- **Scale Factor:** Calibration scale factor, $\beta(\theta_c)/\text{counts}$. Refer to User's Guide for derivation.
- **Output:** Measured signal output of the scattering meter.
- **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

Certificate no: Oxygen Optode 4831_334_179904
 FoilID: 1206EM

Product: Oxygen Optode 4831
 Serial no: 334
 Calibration date: 10.12.2020

2-POINT RECALIBRATION, MULTIPOINT CALIBRATED OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 µM	0 - 120%	-5 to +40°C
Accuracy	< ±3µM or ±2% ¹⁾	±2% ²⁾	±0.03°C ³⁾
Resolution	< 1 µM	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

- 1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter
 2) Within calibrated range 0 - 120% / 0 - 30°C
 3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	33.61	60.94
Temperature reading (°C)	10.05	21.51
Air Pressure (hPa)	957.00	


Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-2.828391E+00	1.042126E+00

With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.80983E-03	1.17557E-04	2.52472E-06	2.31705E+02	-3.27474E-01	-5.62698E+01	4.54625E+00
PhaseCoef	0.000000E+00	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.679277E+01	-3.148383E-02	3.180070E-06	-4.607407E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						
Firmware Version	5.0.4						

Date: 10.12.2020


 Sr. Repair Technician



SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9537
CALIBRATION DATE: 29-Jan-19

Slocum Payload CTD PRESSURE CALIBRATION DATA
1450 psia S/N 11151147

COEFFICIENTS:

PA0 = 2.353944e-001	PTCA0 = 5.238874e+005
PA1 = 4.550810e-003	PTCA1 = 3.600658e+000
PA2 = -1.271461e-011	PTCA2 = -5.638615e-002
PTEMPA0 = -6.574419e+001	PTCB0 = 2.503425e+001
PTEMPA1 = 5.181534e-002	PTCB1 = -3.500000e-004
PTEMPA2 = -4.337839e-007	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.68	527111.8	1708.1	14.68	-0.00	32.50	1927	527141.80
301.70	590182.8	1709.6	301.73	0.00	29.00	1857	527161.10
588.73	653262.9	1710.0	588.73	-0.00	24.00	1758	527140.80
875.83	716385.3	1710.4	875.82	-0.00	18.50	1649	527137.50
1163.00	779550.4	1710.7	1163.00	-0.00	15.00	1579	527131.60
1450.09	842720.5	1711.0	1450.10	0.00	4.50	1371	527108.60
1162.95	779538.1	1710.9	1162.94	-0.00	1.00	1303	527094.00
875.83	716386.8	1710.2	875.82	-0.00			
588.85	653290.4	1710.3	588.85	0.00			
301.68	590173.3	1710.2	301.69	0.00			
14.69	527109.2	1709.9	14.67	-0.00			

	TEMPERATURE (°C)	SPAN
	-5.00	25.04
	35.00	25.02

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$


$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 29-Jan-19 0.00





Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9537
CALIBRATION DATE: 12-Nov-20

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.729775e-004
a1 = 3.232566e-004
a2 = -5.802654e-006
a3 = 2.369666e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	577730.2	0.9999	-0.0001
4.4999	494041.3	4.5001	0.0002
15.0000	315074.1	14.9998	-0.0002
18.5000	272899.3	18.5000	-0.0000
23.9999	219045.0	24.0000	0.0001
29.0000	180469.5	29.0001	0.0001
32.5001	158114.1	32.5000	-0.0001

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

$$\text{Residual (°C)} = \text{instrument temperature} - \text{bath temperature}$$





SEA-BIRD
SCIENTIFIC

Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9537
CALIBRATION DATE: 12-Nov-20

Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.919848e-001
h = 1.343852e-001
i = -1.250621e-004
j = 2.786896e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 2.7588e-007

BATH TEMP (°C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2718.27	0.00000	0.00000
1.0000	34.6935	2.96647	5424.49	2.96647	0.00000
4.4999	34.6739	3.27260	5629.58	3.27261	0.00001
15.0000	34.6330	4.25150	6239.49	4.25147	-0.00003
18.5000	34.6240	4.59560	6439.91	4.59557	-0.00003
23.9999	34.6137	5.15180	6751.03	5.15186	0.00006
29.0000	34.6075	5.67196	7029.14	5.67194	-0.00001
32.5001	34.6015	6.04277	7220.70	6.04276	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

